

FUME HOOD SPECIFICATION

SECTION 11610.7

'PRO' FUME HOODS

ADDENDUM TO ADD COMBINATION SASH

MOTT MANUFACTURING LTD.

SIGMA SYSTEM™

29 April 2003

SECTION 11610.7 – Combination sash option

- 1 Combination sash option.**
- .1 Standard sash shall be deleted and replaced with the combination sash.
- .2 Sash shall be a combination vertical / Horizontal sash.
- .3 Vertically rising frame shall be of stainless steel construction grade 316 or better.
- .4 Horizontal sliding panels shall be 6mm(1/4") laminated safety glass with integral ground-in pulls. Sides of horizontal panels shall be protected with snap-on plastic edge guard.
- .5 Horizontal sliding panels shall ride on rollers in an extruded aluminum bottom and top track with positive locking system to prevent inadvertent removal.
- .6 Sash counterbalance mechanism shall be either chain or cable as required in main hood specification.
- .7 Fume hoods shall also be equipped with a fixed panel located nominally 1" (25mm) behind the sash plane, covering the space between the interior roof of the fume hood and the upper edge of the sash frame. 1" (25mm) space shall be left open to provide downwardly vectored by-pass air, thus reducing dead space behind closed sash.